

CERTIFICATE OF CALIBRATION

Customer Name	FHC Panavally	Certificate No.	FHCPY006
CUSTOMER Address	Panavally, Kerala 688526	Date of Calibration	12/04/2022
		Valid Upto	11/04/2023
		Date of Issue	13/04/2022

DUC* Details			
DUC* Name	REFRIGERATOR	S No.	..
Make/Model No.	WHIRLPOOL	Location	LAB
ID No.	..	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	THERMO HYGROMETER	ELECTRICAL SAFETY ANALYZER	INFRARED THERMOMETER
Make	HTC-2	FLUKE	FLUKE
Model/SR No.	...	ESA 612/2959012	59 MAX/49958584WS
Equipment Id		CHPL/MT-KL/ESA/002	CHPL/MT-KL/IR/001
Calibration certificate No	UCSL/07-21/201-01	ET/220124A/01	UCSL/07-21/201_07
Calibration Validity	20/07/2022	25/01/2023	20/07/2022
Calibration By	UNIQUE CALIBRATION SOLUTIONS	STIC	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET VALUE(°C)	MEASURED VALUE (°C)	TOLERANCE	STATUS
1	5	5	±5%	PASS
2	5	5	±5%	PASS
3	5	5	±5%	PASS

CALIBRATION STATUS	PASS	✓	FAIL
---------------------------	------	---	------

REMARKS:- *DUC – DEVICE UNDER CALIBRATION

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
JAYADEV K S



Issued and Approved
By
PREJITH T B
TECHNICAL
MANAGER

CERTIFICATE OF CALIBRATION

Customer Name	FHC Panavally	Certificate No.	FHCPY006
Address	Panavally, Kerala 688526	Date of Calibration	12/04/2022
		Valid Upto	11/04/2023
		Date of Issue	13/04/2022

DUC* Details					
DUC* Name	Make/Model No	S No.	ASSET ID	LOCATION	Visual Inspection
REFRIGERATOR	WHIRLPOOL	LAB	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	ELECTRICAL SAFETY ANALYZER	THERMO HYGROMETER
Make	FLUKE	HTC-2
Model/SR No.	ESA 612/2959012	...
Equipment Id	CHPL/MT-KL/ESA/002	
Calibration certificate No	ET/220124A/01	UCSL/07-21/201-01
Calibration Validity	25/01/2023	20/07/2022
Calibration By	STIC	UNIQUE CALIBRATION SOLUTIONS

**STANDARD USED IS IEC 60601-1

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229.4	250	225	V	PASS
2	MAIN VOLTAGE -LIVE TO EARTH	228.4	240	220	V	PASS
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.2	3	0	V	PASS
4	EQUIPMENT CURRENT	0.3	16	0.1	mA	PASS
5	PROTECTIVE EARTH RESISTANCE	0.3	2	0	Ω	PASS
6	EARTH LEAKAGE CURRENT - NC	0.6	1	0.5	mA	PASS
7	EARTH LEAKAGE CURRENT - OC	0.7	1	0.5	mA	PASS
8	ENCLOSURE LEAKAGE CURRENT -NC	0.4	0.5	0.1	mA	PASS
9	ENCLOSURE LEAKAGE CURRENT -OC	0.3	0.5	0.1	mA	PASS
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	MΩ	PASS
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	MΩ	PASS
	3.MAINS TO APPLIED PART	NA	100	0	MΩ	PASS
	4.MAINS TO NEUTRAL	OL	100	0	MΩ	PASS
	5.APPLIED PART TO NEUTRAL	NA	100	0	MΩ	PASS

*NOTE (NC -NORMAL CONDITION, OC-OPEN CONDTION). REMARKS- *DUC- DEVICE UNDER CALIBRATION

CALIBRATED BY
JAYADEV K S

ISSUED AND APPROVED BY
PREJITH T B
TECHNICAL MANAGER