

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	FAMILY HEALTH CENTRE PANAVALLY		
ADDRESS	PANAVALLY , POOCHAKKAL , ALAPPUZHZA , KERALA 688526	DATE OF CALIBRATION	20/05/2022
CERTIFICATE NO:	CHPL/PC/FHCP/004	VALID UP TO	19/05/2023
		DATE OF ISSUE	23/05/2022

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S No.	
Make/Model No.		Visual Inspection	OK
LOCATION:	LAB	RANGE	10 µL

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	HTC-2/-
Model/SR No.	MAB182/39736
Calibration Validity	18/02/2023	20/07/2022
Calibration By	UNIQUE	UNIQUE

ENVIRONMENTAL CONDITION			
Temperature	25.6°C	Relative Humidity	40.1%

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	10	10.00	10.01	10.02	9.99	9.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)
PREJITH T B

