

**CERTIFICATE OF CALIBRATION**

CUSTOMER NAME	<b>FAMILY HEALTH CENTRE PANAVALLY</b>		
ADDRESS	PANAVALLY , POOCHAKKAL , ALAPPUZHZA , KERALA 688526	DATE OF CALIBRATION	20/05/2022
CERTIFICATE NO:	CHPL/PC/FHCP/001	VALID UP TO	19/05/2023
		DATE OF ISSUE	23/05/2022

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S No.	PW01285
Make/Model No.	THERMO SCIENTIFIC FINNPIPETTE F2	Visual Inspection	OK
LOCATION:	LAB	RANGE	1000 µL

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER	
Make	WENSAR	HTC-2/-	
Model/SR No.	MAB182/39736	....	
Calibration Validity	18/02/2023	20/07/2022	
Calibration By	UNIQUE	UNIQUE	

ENVIRONMENTAL CONDITION			
Temperature	25.6°C	Relative Humidity	40.1%

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	1000	1000.00	1000.01	1000.02	999.99	999.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE)



Approved By

(QM)

PREJITH T B



**CERTIFICATE OF CALIBRATION**

CUSTOMER NAME	<b>FAMILY HEALTH CENTRE PANAVALLY</b>		
ADDRESS	<b>PANAVALLY , POOCHAKKAL , ALAPPUZHZA , KERALA 688526</b>	DATE OF CALIBRATION	20/05/2022
CERTIFICATE NO:	<b>CHPL/PC/FHCP/003</b>	VALID UP TO	19/05/2023
		DATE OF ISSUE	23/05/2022

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S No.	OW08995
Make/Model No.	<b>THERMO SCIENTIFIC FINNPIPETTE F2</b>	Visual Inspection	OK
LOCATION:	LAB	RANGE	1000 µL

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	HTC-2/-
Model/SR No.	MAB182/39736	....
Calibration Validity	18/02/2023	20/07/2022
Calibration By	UNIQUE	UNIQUE

ENVIRONMENTAL CONDITION			
Temperature	25.6°C	Relative Humidity	40.1%

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	1000	1000.00	1000.01	1000.02	999.99	999.98	PASS

REMARKS:-

<ul style="list-style-type: none"> <li>❖ This result of calibration refers to only to the particular items submitted for calibration.</li> <li>❖ The above result is valid at the time of and under the stated conditions measurement.</li> <li>❖ Calibration certificate issued for weight &amp; measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.</li> </ul>
---

Calibrated By

(SE)



Approved By

(QM)

PREJITH T B

