

CERTIFICATE OF CALIBRATION

| | | | |
|-----------------|--|---------------------|------------|
| CUSTOMER NAME | FAMILY HEALTH CENTRE PANAVALLY | | |
| ADDRESS | PANAVALLY , POOCHAKKAL , ALAPPUZHZA , KERALA 688526 | DATE OF CALIBRATION | 20/05/2022 |
| CERTIFICATE NO: | CHPL/PC/FHCP/006 | VALID UP TO | 19/05/2023 |
| | | DATE OF ISSUE | 23/05/2022 |

| INSTRUMENT DETAILS | | | |
|--------------------|---------------|-------------------|---------------|
| Instrument Name | MICRO PIPETTE | S No. | 061700416 |
| Make/Model No. | LABOPETTE | Visual Inspection | OK |
| LOCATION: | LAB | RANGE | 100 – 1000 µL |

| DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED | | |
|--|--------------------|-------------------|
| Equipment Name | ANALYTICAL BALANCE | THERMO HYGROMETER |
| Make | WENSAR | HTC-2/- |
| Model/SR No. | MAB182/39736 | |
| Calibration Validity | 18/02/2023 | 20/07/2022 |
| Calibration By | UNIQUE | UNIQUE |

| ENVIRONMENTAL CONDITION | | | |
|-------------------------|--------|-------------------|-------|
| Temperature | 25.6°C | Relative Humidity | 40.1% |

| Calibration Results | | | | | | | |
|---------------------|---------------------|-----------------------|---------|---------|--------|--------|--------|
| SL NO | SET READING (µl) | MEASURED READING (µl) | | | | | STATUS |
| | | I | II | III | IV | V | |
| 1 | 100 | 100.00 | 100.01 | 100.02 | 99.99 | 99.98 | PASS |
| 2 | 500 | 500.00 | 500.01 | 500.02 | 499.99 | 499.98 | PASS |
| 3 | 1000 | 1000.00 | 1000.01 | 1000.02 | 999.99 | 999.98 | PASS |

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE) 

Approved By

(QM)
PREJITH T B

