



PC-1033

112th IAMM EQAS Microbiology: Bacteriology/ Serology

CMC MICRO EQAS

Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: egas@cmcvellore.ac.in, Twitter: @microeqas, Phone: +91-416-2282588



NOVEMBER 2022

112th EQAS EVALUATION REPORT

MEMBER ID:

M | 1 | 2 | 0 | 0

Marks Obtained: 6/8 (75%)

NOVEMBER 2022 / SEROLOGY

Please refer the attached evaluation format/answer template for details on the criteria for evaluation.

Parameter	Your Interpretation		Intended Result				Max Marks		Your Score
	Correct	Incorrect	STO	STH	Interpretation	STO	STH		
SE1	Widal*	✓	Negative	Negative	Not Suggestive of Enteric Fever / Typhoid Fever	2	2	4	

*Negative at 1:20 dilution

Parameter	Your Result	Intended result	Correct	Incorrect	Max Marks	Your Score
SE2	TPHA					
	Syphilis ELISA					
NOT DONE						

SE3: Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination
Peer group (n) = 207

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score
SE3 CRP	Positive	Not reported	Positive	31.6832	20.2228	5 to 240	1.7570	Not Applicable	2	2


Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 1 2 0 0

SE1	SE2	SE3	Marks obtained
4	0	2	6
4	2	2	Maximum marks = 8




Dr. Rani Diana Sahni
Scientific Co-ordinator

Report Dispatch Date: 15.02.23



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator

Dr. V. Balaji
Medical Officer
TD-500 (FIMS)
ADILABAD.

***** End of Report *****