



Multitech Technofab Enterprises



475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email:

multitech.technofab@gmail.com, mtedelhi2020@gmail.com, Web. : www.multitechtechnofab.com

CC3225

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/11896

ULR : CC32252300001260F

Certificate Issued on: 18-02-2023

Job / Identification No. : MTE/CAL/JOB/14547

Date of Calibration : 17-02-2023

Item Receipt Date: 16-02-2023

Recommended Date of Next Calibration: 17-02-2024

Calibrated For	JAVIX LIFE PVT. LTD. (Porta Clinic) B-702, Pacific Golf Estate, Sahastradhara Road, Dehradun.	Reference : Order thru Telephonic P. O. No. :	
Discipline	Electro-Technical	Group	Time & Frequency

Certificate for : Digital Timer

Item Location : Lab

Make / Model	Serial / I.D. No.	Inst. Range	Least Count	Unit	Item Condition
NA / DT-109	NA / JVX/PC/DT/2022/02	0 min to 30 min	1	min	Working
Environmental Condition		23.8 ± 0.5 °C		52 ± 3% RH	

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Valid up to
Time Interval Meter	ZEAL / ZMTIM	AACPL/21357F	AA Calibration Lab Pvt. Ltd.	20-10-2023
Calibration Procedure	Lab Procedure : MTE-2/ET/2.19		STD Procedure : NIST 960-12	

RESULTS: In-Lab Calibration

All Readings are in min, Specified Otherwise

UUC READING	CORRECTED VALUE OF STANDARD (AVERAGE) (H:M:S)	ERROR in s	UNCERTAINTY [^]
1	00:00:59.820	0.180	± 0.58
5	00:04:58.600	1.400	± 0.58
15	00:14:57.554	2.446	± 0.58
30	00:29:54.852	5.148	± 0.58

[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.

Calibrated By Ankur Chaturvedi
Technical Executive

For and on behalf of
Multitech Technofab Enterprises

Laboratory Head



Multitech Technofab Enterprises



475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email:

multitech.technofab@gmail.com, mtedelhi2020@gmail.com, Web. : www.multitechtechnofab.com

CC3225

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/11896

ULR : CC322523000001260F

Certificate Issued on: 18-02-2023

3. Total liability of our organization is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

Calibrated By Ankur Chaturvedi
Technical Executive

For and on behalf of
Multitech Technofab Enterprises

Laboratory Head