

**Date: 28/08/2022**

## Certificate of Calibration

Customer Name : "Javix Life Pvt Ltd (PortaClinic)"  
Model : Bio Chemistry Analyzer Erba CHEM 5X NS  
Serial No. : S220790  
Calibration done Date : 28-08-2022  
Next Calibration due Date : On or Before 28-08-2023  
Lab In charge : Mr. Mahesh Kumar Rai

*This is to certify that above mentioned product has been verified of calibration  
By*

*Checking Lamp Voltage  
Checking gain of Photo meter for all filters  
Checking sipping volume of peristaltic pump*

*Followed by calibration of all the three types of chemistries ( picked up CHOLESTROL .  
CREATININE AND SGPT ) and QC run with satisfactory values of QC.*

Calibration at site performed by **Mohit Kumar**.



**(Mohit Kumar)**  
Transasia Biomedicals Ltd.