

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Kathirampatti Lab No 13057

Constituent Group Chemistry III Date of Result Entered : 20/11/2022

PT item Lyophilized human serum based Date of Report Published : 06/12/2022

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	en.	U
						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	82.95	8.62	7.15	84.8 mg/dL	0.26	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1843	6.98	7.13	0.50	7.2 mg/dL	0.44	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1767	112.60	5.74	6.46	109.8 mg/dL	-0.43	0.31

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Kathirampatti Kathirampatti Phc Kathirampatti TAMILNADU635653

Panela Christudoss

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******