



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :
www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/16503

ULR :
CC322523000003543F

Calibration Issued on: 15-03-2023

Job / Identification No. : MTE/CAL/JOB/18409

Date of Calibration : 11-03-2023

Item Receipt Date: 10-03-2023

Recommended Date of Next Calibration: 11-03-2024

Calibrated For

JAVIX LIFE PVT. LTD.
(Porta Clinic) B-702, Pacific Golf Estate, Sahastradhara Road,
Dehradun.

Reference : Order thru Telephonic
P. O. No. :

Discipline

Mechanical

Group

Acceleration & Speed

Certificate for : Centrifuge

Item Location : NA

Make / Model

Serial / I.D. No.

Inst. Range

Least Count

Unit

Item Condition

NA / Labo

NA /
JVX/PC/CENT/2022/01

1000 to 4000

0.0

RPM

Working

Environmental Condition

23.45 ± 0.5 °C

51 ± 3% RH

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard

Make / Model

Certificate No.

Calibration Agency

Valid up to

Digital Tachometer(Non contact)

Lutron / DT-2234C

BCL/2023/0166F

Bhagavati Calibration Laboratory

17-01-2024

Calibration Procedure

Lab Procedure : MTE-2/RPM/2.50

STD Procedure : NA

RESULTS: On-Site Calibration

All Readings are in RPM, Specified Otherwise.

U.U.C. READING (SET AT)

AVERAGE

UNCERTAINTY[^] in (%)

Lower

1006.0

± 2.31

Middle

2511.0

± 2.31

High speed

3999.0

± 2.31

[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

For and on behalf of
Multitech Technofab Enterprises

Calibrated By Sandesh Srivastava
Technical Executive

Laboratory Head



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :
www.multitechtechnofab.com

MULTI TECH
Technofab Enterprises

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/16503

ULR :
CC322523000003543F

Calibration Issued on: 15-03-2023

1. This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
2. The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
3. Total liability of our organization is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

For and on behalf of
Multitech Technofab Enterprises

Calibrated By Sandesh Srivastava
Technical Executive

Laboratory Head