



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

<b>Site name:</b>	Sri Sai Healthcare Unit
<b>Address:</b>	Suchitra Circle, Medchal Road, Hyderabad
<b>Province-Postal code</b>	500067
<b>Contact Name</b>	Mr. Shaik Fayaz

### Calibration Certificate

<b>Date</b>	12/04/23 <small>(dd mmmm, yyyy)</small>	<b>Instrument Model</b>	D-10 Hb Analyzer	<b>Instrument Serial Number</b>	DC3B778104
-------------	--	-------------------------	------------------	---------------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

#### High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Measuring cylinder	11.20/2118	21/11/2020
Specified Range@ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute	1.51	

#### Column Holder temperature Reference: Service Manual: Section 7.9

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Fluke 179	43160402	16/11/2022
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C	35.0	

#### Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.021

The next calibration is due on	11/04/2024	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory: