

## **CALIBRATION CERTIFICATE**

HOSPITAL NAME

**TESLA DIAGNOSTICS,** 

**ADDRESS** 

SUCHITRA.

ANALYSER

COBAS c311

**SERIAL NO** 

17G9-17

**INSTALLATION PLACE** 

**BIOCHEMISTRY LABORATORY** 

DATE

26-02-2023

This is to certify that the above mentioned analyzer was calibrated by the following procedure.

- 1. Checked the supply voltage, and mechanical movements.
- Adjusted the Sample Pipettor probe, Reagent Probe, Reagent Loader, and Mixer mechanism
- 3. Cell Blank, Photometer Check.
- 4. Temperature monitor for incubator, and reagent disk
- 5. Pipettor's probe LLD voltage, and Pressure sensor check

All the checks and calibration were done are found OK and system is being used under recommended conditions. Next calibration is due on 25-02-2024.

For Roche Diagnostics India Pvt. Ltd.

T R Siddardha,

Aidands.

Sr. Technical Service Specialist,

Roche Diagnostics India Pvt Ltd.,

Hyderabad.