

Abbott Healthcare Pvt. Ltd. Floor 13-19, Godrej BKC, Plot C-68 G Block, Bandra Kurla Complex Near MCA Club, Bandra East Mumbai - 400051. India Registered Office: 4, Corporate Park, Sion Trombay Road, Mumbai – 400071. India T : +91 22 3816 2000 / 1010
F : +91 22 3816 2400 / 2401
Email : webmaster@abbott.co.in
Site : www.abbottdiagnostics.com
CIN : U24200MH1997PTC104834

Preventive Maintenance & Calibration Certificate

This is to inform that the Abbott Architect ci4100sr fully automated Biochemstry & Immunology Analyzer bearing the serial number c402946 & i1sr62499 Installed at Telangana Diagnostics Hub ,Medak has been performed with the PM & Hardware Calibration of the instrument. The result are within the acceptable limit. The Hardware Calibration & PM was performed on 18th May 2023.

The next Preventive Maintenance and Hardware Calibration due is on 17th May 2024

For Abbott Healthcare Pvt Ltd.



Mahesh.Lokam (Technical support specialist) procedure: 1032 Shutter Test

procedure version: 3 Module no.: 2

Serial no.: i1SR62499

Operator ID: ADMIN

Date: 18.05.2023

Time: 01:17:19

Status: Completed

Results:

Shutter open sensor

Pass

Shutter closed sensor

Pass

Shutter open sensor

Pass

Shutter closed sensor

Pass

Procedure completed.

Printed on: 18.05.2023

1:17:21AM

ARCHITECT

Page: 1 of 1

pocedure: 1121 R1 Pipettor Calibration

procedure version: 10

Module no.: 1

Serial no.: c402946

Operator ID: ADMIN

Date: 18.05.2023

Time: 01:46:12

Status: Completed

Results:

Calibration Point	Theta	z	Carousel
Wash Cup	1		
	·	NA	NA
Reaction Carousel			
	1	NA	2
Outer Segment			
outer degriferit	-1	3	8
Inner Segment	-1		10
Ç sam	-1	4	-18

Calibration passed