



Metropolis Healthcare Ltd., EQAS

P.H. MEDICAL CENTRE KANDIVALI (898)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230104 and Sample No 02

Report Date : 19/05/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Bilirubin Total	RANDOX RX imola	0.66	mg/dl	4.28
X Bilirubin Direct	RANDOX RX imola	0.54	mg/dl	12.88
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : RANDOX RX imola

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	84.90	84.45	1.78	2.13
@ Urea	mg/dl	20.70	20.70	-1.48	-0.98
@ Creatinine	mg/dl	0.98	0.98	-0.04	-0.01
X Bilirubin Total	mg/dl	0.66	0.10	4.28	5.21
X Bilirubin Direct	mg/dl	0.54	0.05	12.88	12.71
@ Total Protein	g/dL	6.70	6.90	-0.42	-0.38
@ Albumin	g/dL	4.00	4.10	0.11	0.33
@ Uric Acid	mg/dl	4.30	4.45	-0.59	-0.33
@ Calcium	mg/dl	8.60	8.65	-0.29	-0.47
@ Phosphorous Inorganic	mg/dl	4.50	--	-0.32	-0.13
@ Cholesterol	mg/dl	174.50	170.25	0.98	1.33
@ Triglycerides	mg/dl	138.70	146.35	0.17	0.80
@ HDL	mg/dl	30.50	--	-1.93	-1.40
@ ALT	U/L	7.00	7.00	-0.92	-0.40
@ AST	U/L	19.20	19.20	1.04	1.74
@ ALP	U/L	47.00	59.50	-1.13	-1.76
@ GGT	U/L	43.90	--	0.74	0.57

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	0
Evaluated Parameters	17
Outlier Parameters (X)	2
EQAS Score Biochemistry	88.24 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070