

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.38	2.38	0.22	0.42
@ Total Leucocytes (WBC) count	/c.mm	8450.00	7541.11	1.85	0.99
@ Haemoglobin	g/dL	6.70	6.84	-0.62	-0.31
@ MCV (Mean Corpuscular Volume)	fL	110.10	110.71	-0.20	0.12
@ MCH (Mean Corpuscular Hb)	pg	28.30	28.84	-0.47	-0.58
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	25.70	26.17	-0.04	-0.64
@ Platelet Count	1000/ μ L	233.00	227.30	1.06	0.71
@ Hematocrit	%	26.20	26.35	-0.09	0.50
@ RDW CV%	%	19.30	19.24	0.45	0.37

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070