



**HORIBA India Private Ltd.**

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2023-2024/00297

**29<sup>th</sup> April 2023.**

**CALIBRATION CERTIFICATE**

This is to certify that the Hematology Analyzer **ABX Yumizen H550** bearing serial number: **202YAXH03674** installed at **P. H. Medical Centre, Kandivali - Mumbai** Calibrated on **29<sup>th</sup> April 2023.**

Calibrator : ABX MINOCAL  
Lot No. : CX482  
Expiry Date : 05<sup>th</sup> June 2023.

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

**Next calibration cycle is due on 28<sup>th</sup> April 2024.**



**Shrish Dixit**

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**