



Metropolis Healthcare Ltd., EQAS

HEER CLINICAL LABORATORY (887)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230104 and Sample No 04

Report Date : 19/05/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Mindray BC 760

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.25	--	-0.44	-0.52
@ Total Leucocytes (WBC) count	/c.mm	8800.00	--	0.22	-0.20
@ Haemoglobin	g/dL	7.20	--	-0.31	-0.33
@ MCV (Mean Corpuscular Volume)	fL	142.00	--	1.18	1.35
@ MCH (Mean Corpuscular Hb)	pg	32.10	--	0.02	0.00
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	22.60	--	-0.64	-0.74
@ Platelet Count	1000/ μ L	221.00	--	-0.55	-0.48
@ Hematocrit	%	32.00	--	0.57	0.58
@ RDW CV%	%	19.40	--	0.69	0.71

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070