



Metropolis Healthcare Ltd., EQAS

HEER CLINICAL LABORATORY (887)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230104 and Sample No 01

Report Date : 19/05/2023

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : DocUreader 2Pro

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5.5	5,5,5,6
@ Glucose (urine)	-	4+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,Tr
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	Negative	Neg
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 Ⓞ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirool Road, Kurla (W),
 Mumbai - 400070