



Metropolis Healthcare Ltd., EQAS

HEER CLINICAL LABORATORY (887)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230104 and Sample No 01

Report Date : 19/05/2023

Analyte	Instrument	Result Value	Standard Unit
X C-Reactive Protein Qualitative	Mindray BS-240	Normal	mg/L
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification:

Corrective Action:

Reviewed by:

Dated:

Instrument : Mindray BS-240


Analyte	Standard Unit	Result Value	Accepted Value
X C-Reactive Protein Qualitative	mg/L	Normal	Abnormal
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
@ ASO Qualitative	IU/mL	Normal	Normal

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	NR
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	13
Not Evaluated Parameters	0
Evaluated Parameters	13
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	92.31 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070