

Date: 14/10/2022

Certificate of Calibration

Customer Name: Sankar Diagnostic Centre , Agra

Model: Fully Automated Bio Chemistry Analyzer EM-200

Serial No.: B160410

Calibration done date: 14/10/2022

Next Calibration due Date: On or Before: 13/10/2023


Lab Incharge: Dr. Meghna Singh

*This is to certify that above mentioned product has been verified of calibration
By*

*Checking Lamp Voltage
Checking gain of Photo meter for all Filters
Checking Cell blanks value
Checking laundry section
Checking movement of probe*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of
QC*

Calibration at site performed by Pushpendra Singh Bhadauria


Pushpendra Singh Bhadauria
Sr. Service Engineer
Transasia Biomedicals Ltd.
Agra