

Date: 14/10/2022

Certificate of Calibration

Customer Name: Sankar Diagnostic Centre, Agra

Model: Semi Automated Bio Chemistry Analyser Erba Chem7

Serial No.: N120232

Calibration done date 14/10/2022

Next Calibration due Date: On or Before 13/10/2023

Lab In charge: Dr. Meghna Singh

*This is to certify that above mentioned product has been verified of calibration
By*

*Checking Lamp Voltage
Checking gain of Photo meter for all Filters
Checking sipping volume of peristaltic pump*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of
QC*

Calibration at site performed by Sharad Srivastava

Pushpendra Singh Bhadauria
Sr. Service Engineer
Transasia Biomedicals Ltd.
Agra