



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JANUARY 2023



Lab Name **TELANGANA DIAGNOSTIC HUB** Lab No **18032**  
Constituent Group **Chemistry I** Date of Result Entered : **25/01/2023**  
PT item **Lyophilized Serum** Date of Report Published : **09/02/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	157	174.21	4.68	8.15	175 mg/dL	0.10	1.30
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	588	31.57	10.65	3.36	14 mg/dL	-5.23	0.28
3	CREATININE	Enzymatic Colorimetric	Any Analyser (Automation / Semi Automation)	138	1.39	10.91	0.15	1.23 mg/dL	-1.05	0.03
4	T.BILIRUBIN	Others (DPD, Vanadate Oxidation)	Any Analyser (Automation / Semi Automation)	205	3.20	12.17	0.39	4.3 mg/dL	2.82	0.05
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	956	5.25	9.77	0.51	5.1 g/dL	-0.29	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	674	3.31	8.11	0.27	3.2 g/dL	-0.41	0.02
7	CALCIUM	OCPC (O - Cresolphthalein Compleazone)	Any Analyser (Automation / Semi Automation)	177	9.32	8.24	0.77	8.9 mg/dL	-0.55	0.12
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	939	4.83	16.02	0.77	4 mg/dL	-1.07	0.05
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	974	113.44	5.89	6.68	118 mg/dL	0.68	0.43
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	935	109.13	7.82	8.54	120 mg/dL	1.27	0.56
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	780	30.17	20.45	6.17	26.7 mg/dL	-0.56	0.44
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	969	132.87	2.93	3.88	139.3 mmol/L	1.71	0.25
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	997	3.42	5.15	0.18	3.52 mmol/L	0.57	0.01
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	809	97.51	4.13	4.03	106.3 mmol/L	2.18	0.28
15	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	967	50.29	10.00	5.03	65 U/L	2.92	0.32
16	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	958	57.78	13.00	7.51	68 U/L	1.36	0.49
17	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	793	101.50	10.81	10.97	91 U/L	-0.96	0.78
18	AMYLASE	Enzymatic Colorimetric / G7PNP Blocked	Any Analyser (Automation / Semi Automation)	154	44.10	17.79	7.84	56 U/L	1.52	1.26

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
TELANGANA DIAGNOSTIC HUB  
GOVERNMENT GENERAL HOSPITAL

2/10/23, 1:22 PM

External Quality Assurance Scheme - Print Monthly Summary

NIZAMABAD

TELANGANA503001

**Coordinator Contact Details:**  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
**Christian Medical College, Vellore**

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***



TELANGANA DIAGNOSTICS HUB,  
NIZAMABAD

Form: TD/QSP/08-EQCAR

TITLE

EQAS CORRECTIVE ACTION FORM

Issue No. 01

Page 1 of 1

EQAS Details	CMC-EQAS
Analyte:	Chemistry - I
Month:	January - 2023
Date Sample Tested:	24-01-2023 & 25-01-2023

SPECIMEN HANDLING		
Were specimens received in an acceptable condition?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were specimens stored according to the instructions on the result forms?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the samples hemolyzed?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Were samples tested within the time allowed for sample stability?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
If applicable, were the samples reconstituted correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Notes: CMC-EQAS sample is reconstituted correctly with 2ml of distilled water and run 2 times consecutively		
CLERICAL ERRORS		
Were the results transcribed onto the result forms correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results transcribed from the result forms to the website correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded on the correct result form?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Was the correct instrument/reagent/kit selected?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded in the correct units? (except for Urea)	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results on your evaluation the same as the results you reported?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Notes: we have crosschecked all the methods and then entered results and then submitted.		
QUALITY CONTROL		
Were quality control materials within the acceptable range on the date of PT testing? (Verify the quality control acceptable range in use.) - (except for calcium)	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Is there any indication of trending or shifting of the control results?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Notes: IQC is within the range for all analytes except for calcium.		
CALIBRATION		
Were there any problems with the most recent calibration? - for calcium.	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
When was the last calibration performed?	As per the manufacturer's instructions	
How often is a calibration performed?	As per the manufacturer's instructions	
When was the last calibration verification performed?	Routinely done whenever calibration is done.	
Notes: calibration is done for every analyte as per the manufacturer's instructions and verified and then accepted.		
INSTRUMENT - Minor Technical issues. (- Siemens Atellica) Analyser.		

PREPARED BY : LAB MANAGER : SIRISHA	APPROVED & ISSUED BY: LAB HEAD: Dr. NADIPALLY DIVYA
REVIEWED BY CONSULTANT BIOCHEMIST Dr. M. SWATHI	

CONTROLLED COPY



**TELANGANA DIAGNOSTICS HUB,  
NIZAMABAD**

**Form: TD/QSP/08-EQCAR**

**TITLE EQAS CORRECTIVE ACTION FORM**

Issue No. 01  
Page 1 of 1

Were instrument problems noted the day the samples were tested? - <i>Minor breakdowns</i>	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Has there been any recent maintenance on the analyzer?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Have you contacted your analyzer manufacturer for assistance?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>

Notes: *Minor breakdowns like reagent probe alignment issues to soft waste consumption issues - contacted manufacturers and rectified!*

REAGENTS		
Were the reagents stored properly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the reagents expired or was the open vial stability exceeded?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Have there been any changes in reagent manufacturer or formulation?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>

Notes: *All the reagents were stored properly.*

TESTING PERSONNEL		
Date of last competency assessment for testing personnel <i>10-10-2022</i>	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Review assay procedure and proficiency test sample preparation instructions with testing personnel to ensure that instructions were followed	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Review with testing personnel how samples were loaded to rule out misidentification or transposition of samples.	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>

Notes: *The technicians were trained and the faculty at T-Hub supervise the routine work, calibration and IQC and take appropriate corrective actions. EQAs sample is run routinely as patient samples.*

Corrective Action: *→ The CME sample (Jan) is aliquotted after run and stored at -20°C. This sample is re-run for all the analytes, crossed 2SD, on 13-02-2023 and the results used in acceptable range.*

- ① AST value is 60 and in range ( $50.29 \pm 2 \times 5.03$ )
- ② Urea value is wrongly entered as BUN value and re-run value is 30.8 and acceptable ( $31.57 - \text{Mean} \pm 2 \times 3.36$ )
- ③ T. Bilirubin value, re-run is 3.8 and in range ( $3.20 \pm 2 \times 0.39$ )
- ④ chloride re-run value is 104 and in range ( $97.51 \pm 2 \times 4.09$ )

Person Performing Investigation: *Ms. Shiva (L.T-Biochemistry)* Date: *24 & 25 Jan-2023*  
 Consultant Biochemist *Dr. M. Swathi* Date: *16-02-2023*

PREPARED BY : LAB MANAGER : SIRISHA	APPROVED & ISSUED BY: LAB HEAD: Dr. NADIPALLY DIVYA
REVIEWED BY CONSULTANT BIOCHEMIST  Dr. M. SWATHI <i>Swathi</i>	<i>N. Divya</i>



<b>TELANGANA DIAGNOSTICS HUB, NIZAMABAD</b>		<b>Form: TD/QSP/08-EQCAR</b>
<b>TITLE</b>	<b>EQAS CORRECTIVE ACTION FORM</b>	Issue No. 01 Page 1 of 1

**INVESTIGATION SUMMARY: ROOT CAUSE**

<u>Pre-analytic Phase of Testing</u>	<u>Analytic Phase of Testing</u>	<u>Post-Analytic Phase of Testing</u>
<input type="checkbox"/> PROBLEM WITH PT SAMPLE <input type="checkbox"/> SAMPLE PROCESSING <input type="checkbox"/> DATA ENTRY <input type="checkbox"/> OTHER (SPECIFY): <u>-No-</u>	<input type="checkbox"/> METHODOLOGICAL PROBLEM <input type="checkbox"/> TECHNICAL PROBLEM <input type="checkbox"/> REAGENT PROBLEM <input type="checkbox"/> CALIBRATOR PROBLEM <input type="checkbox"/> OTHER (SPECIFY): <u>-No-</u>	<input checked="" type="checkbox"/> CLERICAL ERROR <input type="checkbox"/> REPORTING PROBLEM <input type="checkbox"/> NO EXPLANATION AFTER INVESTIGATION. OTHER (SPECIFY): <u>For Urea</u>

**PREVENTION**

Preventive action proposed

- Here after, maintenance schedule and calibration schedule is followed as per S.O.P. IQC is reviewed daily by faculty.  
 - In view of frequent minor breakdowns, preventive maintenance of instrument is recommended & contacted the manufacturer.

Preventive action Plan

- Urea value is enabled for CMC-EQAS (instead of BUN)  
 - All parameters IQC is maintained ~~for~~ and supervised by faculty of Biochemistry. If any outliers/trends/shifts, corrective action is taken accordingly.

Responsibility

The faculty of Biochemistry in T-Hub supervise the routine work under the guidance of consultant Biochemist and Lab Director.

Date <u>13-02-2023</u> <u>CR-Sunt</u>	Testing Personnel <u>Ms. Shiva (L.T)</u>
Date <u>16-02-2023</u>	Department Technical In charge <u>Dr. B. Sandeep</u> - <u>Dr. M. Swathi</u>

Method - Swathi  
Dr. M. Swathi  
 Consultant Biochemist

<b>PREPARED BY : LAB MANAGER : SIRISHA</b>	<b>APPROVED &amp; ISSUED BY:</b>
<b>REVIEWED BY CONSULTANT BIOCHEMIST</b>	<b>LAB HEAD: Dr. NADIPALLY DIVYA</b>
<b>Dr. M. SWATHI</b>	



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - FEBRUARY 2023



Lab Name	TELANGANA DIAGNOSTIC HUB	Lab No	18032
Constituent Group	Chemistry I	Date of Result Entered :	20/02/2023
PT item	Lyophilized human serum based	Date of Report Published :	08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	182	120.06	5.58	6.70	120 mg/dL	-0.01	0.99
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	648	129.71	8.40	10.90	128.7 mg/dL	-0.09	0.86
3	CREATININE	Enzymatic Colorimetric	Any Analyser (Automation / Semi Automation)	151	5.17	6.52	0.34	4.95 mg/dL	-0.65	0.05
4	T.BILIRUBIN	Others ( DPD, Vanadate Oxidation )	Any Analyser (Automation / Semi Automation)	211	1.48	14.72	0.22	1.8 mg/dL	1.47	0.03
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	1100	4.63	9.80	0.45	4.1 g/dL	-1.17	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	730	2.91	8.69	0.25	2.7 g/dL	-0.83	0.02
7	CALCIUM	OCPC ( O - Cresolphthalein Complezone )	Any Analyser (Automation / Semi Automation)	195	10.23	8.50	0.87	10.4 mg/dL	0.20	0.12
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	1011	6.92	10.14	0.70	6.2 mg/dL	-1.03	0.04
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	1046	98.66	7.43	7.33	100 mg/dL	0.18	0.45
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	988	87.48	8.82	7.72	92 mg/dL	0.59	0.49
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	714	28.16	19.26	5.42	20.3 mg/dL	-1.45	0.41
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1147	120.46	4.12	4.96	126.4 mmol/L	1.20	0.29
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1129	5.00	5.46	0.27	4.86 mmol/L	-0.51	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	921	92.85	4.47	4.15	98.7 mmol/L	1.41	0.27
15	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1021	190.14	15.11	28.72	251 U/L	2.12	1.80
16	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1025	49.30	15.70	7.74	52 U/L	0.35	0.48
17	AMYLASE	Enzymatic Colorimetric / G7PNP Blocked	Any Analyser (Automation / Semi Automation)	187	79.27	16.96	13.45	102 U/L	1.69	1.97

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
TELANGANA DIAGNOSTIC HUB  
GOVERNMENT GENERAL HOSPITAL  
NIZAMABAD  
TELANGANA503001

*Pamela Christudoss*

**Coordinator Contact Details:**  
**Email:** [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
**Contact Number:** 0416-2283102

**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**  
**Data in CMC EQAS reports is confidential**  
**CMC EQAS does not sub contract any components**  
**\*\*\*\*\* End of Report \*\*\*\*\***