



Metropolis Healthcare Ltd., EQAS

RAHA DIAGNOSTICS (824)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220412 and Sample No 03

Report Date : 25/01/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Bilirubin Total	mg/dl	0.44	0.43	0.13	0.13
@ Bilirubin Direct	mg/dl	0.21	0.18	0.36	-1.18
@ ALT	U/L	35.50	37.69	-0.54	-0.69
@ AST	U/L	41.70	49.65	-0.92	-1.10

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	4
Not Evaluated Parameters	0
Evaluated Parameters	4
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirool Road, Kurla (W),

Mumbai - 400070