



Metropolis Healthcare Ltd., EQAS

RAHA DIAGNOSTICS (824)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220412 and Sample No 03

Report Date : 25/01/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Sysmex XN 330

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.19	1.19	-0.21	-0.49
@ Total Leucocytes (WBC) count	/c.mm	1380.00	1410.00	-0.10	0.10
@ Haemoglobin	g/dL	3.10	3.10	0.00	-0.49
@ MCV (Mean Corpuscular Volume)	fL	121.00	123.65	0.22	0.47
@ MCH (Mean Corpuscular Hb)	pg	26.10	26.20	0.06	0.06
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.50	21.15	-0.20	-0.52
@ Platelet Count	1000/ μ L	25.00	31.50	-1.49	-0.74
@ Hematocrit	%	14.40	14.65	-0.39	0.01
@ RDW CV%	%	19.50	20.05	0.00	0.38

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070