



Metropolis Healthcare Ltd., EQAS

RAHA DIAGNOSTICS (824)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220412 and Sample No 01

Report Date : 25/01/2023

Analyte	Instrument	Result Value	Standard Unit
# C-Reactive Protein Qualitative	Erba EM 200	Normal	mg/L
X Dengue IgG	Manual	Positive	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ Delayed Result Entry

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Dengue IgM	RU/ml	Negative	Negative
X Dengue IgG	RU/ml	Positive	Negative

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Accepted Value
# C-Reactive Protein Qualitative	mg/L	Normal	
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
@ ASO Qualitative	IU/mL	Normal	Normal

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	10
Not Evaluated Parameters	1
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	88.89 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070