

LABTECH HEALTH CARE / JPR/455 / 2021-22

Date: -01/12/2022

CALIBRATION CERTIFICATE

This is to certify that the below-mentioned Instrument was calibrated on 03/02/2023 at **BIO 5 LAB, Mangalam Plaza, Door no: 226/5A, MGR Main Rd, Kandhanchavadi, Perungudi, Chennai, Tamil Nadu 600096.** And found working satisfactorily.

Instrument : Electrolyte Analyzer
Name Model : Lablyte
Serial No : LTE19518

Calibration Done on : 01/12/2022
Next Calibration Due on : 01/12/2023

CALIBRATION PARAMETERS:

1. Input Power - Ok
 - i. Ph To N -229 Volts
 - ii. E To Ph -232 Volts
 - iii. N To E - 03 Volts
2. Display Test - Ok
3. Touch Screen Test - Ok
4. Pump Calibration Test -Ok
5. Optical Calibration Test - Ok
6. Printer Test - Ok

Yours Faithfully,

For Labtech Healthcare India Pvt. Ltd.

(Authorized Signatory)