

Date: 25 Jan 2023

CALIBRATION CERTIFICATE

This is to certify that Fully **Automated Haematology Analyser XN-330**, Serial No: 14805
Calibrated at RAHA DIAGNOSTICS, THRISSUR-679563

Calibration Done on: 13 Jan 2023
Calibration Validity: 12 Jan 2023 to 11 Jan 2024
Calibration Due on: 12 Jan 2024 on or Before

Calibration Information:

Calibration Material	Calibrated Date	Lot no	Exp
XN CAL	13 Jan 2023	23602101	2023-01-29

For Sysmex India Pvt. Ltd.,



Authorised Signatory

Calibrator Calibration History

Instrument Nickname: XN-L
Material: XN CAL

Calibration Date: 2023/01/13 17:25:56
Lot No.: 23602101

	WBC	RBC	HGB	HCT	PLT
Target	7.741	4.375	12.33	35.88	254.9
No. 2	8.03	4.40	12.3	35.9	259
No. 3	7.96	4.40	12.3	35.9	248
No. 4	7.84	4.39	12.3	35.9	265
No. 5	7.73	4.39	12.4	35.8	264
No. 6	7.75	4.40	12.4	35.9	253
No. 7	7.84	4.45	12.2	36.3	252
No. 8	7.87	4.38	12.4	35.7	253
No. 9	7.95	4.43	12.4	36.2	257
No. 10	7.76	4.43	12.3	36.2	255
No. 11	7.78	4.45	12.5	36.2	253
Range Value	0.30	0.07	0.3	0.6	17
Max Range	0.60	0.13	0.2	1.0	27
Mean Value	7.851	4.412	12.35	36.00	255.9
Delta Percent (%)	1.40	0.84	0.16	0.33	0.39
Acceptable Limit (%)	2.27	1.25	0.78	2.64	4.16
Service Limit (%)	14.00	4.00	5.00	5.00	10.00
Current Rate (%)	106.0	101.4	99.9	101.6	102.3
New Rate (%)	106.0	101.4	99.9	101.6	102.3