

Calibration Certificate

This is to certify that the Cobas Integra 400 Plus is a fully automated analyzer with **S.No:422088** at **Tenet Medcorp Pvt Ltd**, Visakhapatnam has been calibrated on **15th May 2023**.

The calibration includes:


Adjustments:

1. Checked and adjusted cuvette Rotor.
2. Sample/Reagent probe, pippetors movements.
3. Mixer and their respective home positions.
4. Reagent and incubator temperature.
5. Cuvette transportation system.
6. Fluidics system.

All the above checks and calibration were done and system is being used under recommended conditions.

Next calibration is due on **14th May 2024**.

For Roche Diagnostics India Pvt Ltd.,



Naresh Gutta
Assistant Consultant – Technical
Chennai