



## CALIBRATION CERTIFICATE

<b>NAME OF THE CUSTOMER</b>	<b>BIOLINE LABORATORY. HYDERABAD.</b>
-----------------------------	---

Make	<b>Beckman Coulter</b>
PRODUCT NAME	<b>DxH 800-Haematology Analyser</b>
SERIAL NO.	<b>RBC04011</b>

This is to certify that the **DxH 800 Haematology analyser** has been calibrated with S- cal calibrator (Lot.No-1131162310, expiry-04/29/ 2022)

S.NO.	Procedures	parameter	REMARKS
01	Start up	WBC/RBC/HGB/PLT	PASS
02	Reproducibility	WBC/RBC/HGB/MCV/HCT/PLT	PASS
03	Carry over	WBC/RBC/HGB/PLT	PASS
04	QC	WBC/RBC/HGB/HCT/PLT	PASS
05	Calibration (Lot.no-113162310)	WBC/RBC/HGB/MCV/PLT/MPV	PASS

**Note: Minimum 5 runs have been taken for auto calibration verification**

<b>Date of calibration:04/25/2023</b>	<b>Due date for calibration:04/24/2024</b>
---------------------------------------	--

**For Beckman Coulter India private Limited**

**Suresh Duggana**  
**Sr. Application Manager**