



## **CALIBRATION CERTIFICATE**

Customer name	M/S: Labxperts Technologies	Service request No. & date	A/02/04/03/23&04/03/2023
And	Shop no E 14, ESHF 7, GDA	ULR no.	CC223923000001634F
Address	Market, Patel NAGAR 2nd,	Cert. No.	AT23000001634
	Ghaziabad - 201001	Date of Receipt of DUC	04/03/2023
	Gliaziabad - 201001	Date of calibration	06/03/2023
		Date of issue	07/03/2023
		Suggested due date	05/03/2024

	Instrument Deta	ils	
Instrument name	Centrifuge Machine	Sr. No.	202108271-12
Make	EURO	Location	
Range / Size	0 to 4000 rpm	Model No.	ELECTRA-BL
Least Count	1 rpm	Visual Inspection	OK
I.D. No.			

	Detail of reference standard	s & Major equipments used	
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	/A.C92792		
Certificate No.	QCTS/000246/10		
Calibration Validity	20/10/2023		
Calibration by	QCTS		

Environmental	Temperature	20±2°C Calibration Reference IS: 12508			
I	Condition	Relative Humidity	50±10 %	Calibration Procedure	CD 27
	The second secon			Cantilation Procedure	CP-3/

Calibration Results

Serial No.	Parameter	Standard value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	1790	1700	
02.	2270	1788	
03.	3175	2271 3177	
04.	3580	3582	±1 %rdg
05.	3900	3901	

## Remarks:

(1) Standard equipment use for calibration are traceable to national/international standards.

(2) The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given

Calibrated By ration Engg./TM)

