

To Whom It May Concern

For ISO 15189:2012 and ISO 15189:2014 accredited Laboratories — requirements regarding "Calibration & Verification Procedures" [1]

All In vitro Diagnostics Products which are manufactured and distributed by Roche Diagnostics GmbH and for which a Free-Sales-Certificate is issued, are CE-marked.

The In-Vitro-Diagnostics Directive of the European Union [2A.] which is currently switching to IVD Regulation 2017/746/EU (final timeline: May 26, 2022) [2B.] requires for all CE marked products that the manufacturer assures compliance of the products with the requirements of the mentioned directive or regulation. This means that all processes in development and manufacturing of Roche Diagnostics GmbH products are guided by a Quality Management System. Our Quality Management System is in compliance with the requirements from ISO 13485:2016 [3] and 21 CFR Part 820 [4]. The mentioned regulations and standards require that the production systems and measuring devices used are qualified and the manufacturing and test procedures are validated. This status has to be assured by scheduled maintenance and by regular qualification resp. validation reviews and updates.

All physical quantities, calibrators and controls used in Roche Diagnostic systems are fully traceable to certified standards or reference materials. The performance of all In-vitro diagnostics systems of Roche Diagnostics GmbH at the customer site is assured if regular Quality Control measurements, cleaning and maintenance procedures as described in the instructions for use or service documentation are performed. By having controlled internal procedures and by running the tasks required in the respective user documentation, all In-vitro diagnostics systems of Roche Diagnostics GmbH will be performed as specified during their defined lifetime.

Additional calibration or verification procedures are NOT required by the user in order to assure the specified performance of every system of Roche Diagnostics GmbH. Only if a user deviates from these manufacturer's recommendations, the user have to establish site-specific calibration and verification procedures as part of his accreditation process.



- [1] ISO 15189:2012/ ISO 15189:2014 Medical laboratories Requirements for quality and competence
- A. Directive 98/79/EC of the European Parliament and of the Council of the 27 October 1998 on vitro diagnostics medical devices;
 B. IVD Regulation 2017/746/EU of the European Parliament and of the Council of 5 April 2017 on in vitro diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU
- [3] EN ISO 13485:2016 Medical devices Quality management systems-Requirements for regulatory purposes
- [4] CFR Part 820, Quality System regulations 21 Regulations on medical devices

Mannheim, 10. August 2021

Sincerely,

Roche Diagnostics GmbH

i.V./on behalf of the company

ppa/on behalf of the company

Ralf Ziclenski

Ralf Zielenski Head Q&R Compliance, PRRC RDG Centralised and Point of Care Solutions

Andrea Weber

Manager Global Regulatory Affairs Centralised and Point of Care Solutions

Roche Diagnostics GmbH Sandhofer Straße 116 D-68305 Mannheim

Field Service Report



Case number: CAS-0016640862 Order Number: WO-03246864 | Ser. Report Created Date:

Instrument Details Customer details

Instrument/Module: cobas 6000 core unit Customer Number: 0052615703

Serial Number: B993-05 Customer Name: Bioline Laboratory

Internal Instrument Name: Street Address: 6, 7 & 15, New Nagole,

Zip: 500068 - Hyderabad

Additional Details Contact Name: Mr. Ravi

Contact Phone: +917569727272

Service Activity Code: PM Visit

Purpose of Visit

PM cobas 6000 core unit Preventive Maintanence

Performed Activities

Action Summary: Performed dusting,

Lubricated robotics and checked alignments and found ok.

Run Mechanism check and found ok.

Engineer comments: instrument is ok.

Time Report

Category	Start	End	Hours	Invoice Type
05913616001-Service labour time	13/02/2023 8:30 PM	13/02/2023 9:39 PM	1.15	Contract

Field Service Report



Case number: CAS-0016640862 Order Number: WO-03246864 | Ser. Report Created Date:

Signature

Customer Signature

Mr. Ravi

Roche support Tadiparthi R Siddardha

Sign Here



Roche Signature

The customer acknowledges the service intervention as performed in accordance with Roche recommendations:

- This Service report has been signed by the authorized representative of your organization.
- Discrepancies if any must be intimated within a period of seven (7) days of the receipt of the email, in the absence of which the report will be final and conclusive.

- This report shall be governed by the laws of India and the courts at Mumbai shall have the exclusive jurisdiction to deal with the disputes arising hereunder.

Customer Support Center No: Toll Free No: 1800-123-7599 || Phone No: 044-43900345 ||

(Mon to Sat 8 AM to 8 PM)

e-mail: customersupportcenterIndia@roche.com ||||||||||||

Roche DiaLog: https://dialog.roche.com Service Report Version Number: 1

Field Service Report



Case number: CAS-0015526406 Order Number: WO-01087197 | Ser. Report Created Date: 13/02/2023

Instrument Details Customer details

Instrument/Module: cobas e 601 Customer Number: 0052615703

Serial Number: 31E4-05 Customer Name: Bioline Laboratory

Internal Instrument Name: Street Address: 6, 7 & 15, New Nagole,

Zip: 500068 - Hyderabad

Additional Details Contact Name: Mr.Ravi

Contact Phone: +917569727272

Service Activity Code: PM Visit (Interim)

Purpose of Visit

Preventive maintenance

Performed Activities

Action Summary:

Performed Dusting and replaced pm kit.

Performed lubrication and checked and found ok.

Performed allignments and found ok.

Run SVC and found ok.

Run Qc and Samples and found ok.

Engineer comments: Instrument is ok.

Time Report

Category	Start	End	Hours	Invoice Type
05913616001-Service labour time	13/02/2023 2:30 PM	13/02/2023 7:30 PM	5.00	Contract

Travel

Value	Hours
Travel	0.5

Spare Parts

Part Number	Description	Quantity	Invoice Type
03504620001-KIT MAINTENANCE E601/E602	KIT MAINTENANCE E601/E602	1	Contract

Field Service Report



Case number: CAS-0015526406 Order Number: WO-01087197 | Ser. Report Created Date: 13/02/2023

Signature

Customer Signature Mr.Ravi Roche support Tadiparthi R Siddardha

Sign Here



Roche Signature

The customer acknowledges the service intervention as performed in accordance with Roche recommendations:

- This Service report has been signed by the authorized representative of your organization.
- Discrepancies if any must be intimated within a period of seven (7) days of the receipt of the email, in the absence of which the report will be final and conclusive.
- This report shall be governed by the laws of India and the courts at Mumbai shall have the exclusive jurisdiction to deal with the disputes arising hereunder.

Customer Support Center No: Toll Free No: 1800-123-7599 || Phone No: 044-43900345 ||

(Mon to Sat 8 AM to 8 PM)

e-mail: customersupportcenterIndia@roche.com ||||||||||||

Roche DiaLog: https://dialog.roche.com Service Report Version Number: 1

Field Service Report



Case number: CAS-0016640862 Order Number: WO-03246864 | Ser. Report Created Date:

Instrument Details Customer details

Instrument/Module: cobas 6000 core unit Customer Number: 0052615703

Serial Number: B993-05 Customer Name: Bioline Laboratory

Internal Instrument Name: Street Address: 6, 7 & 15, New Nagole,

Zip: 500068 - Hyderabad

Additional Details Contact Name: Mr. Ravi

Contact Phone: +917569727272

Service Activity Code: PM Visit

Purpose of Visit

PM cobas 6000 core unit Preventive Maintanence

Performed Activities

Action Summary: Performed dusting,

Lubricated robotics and checked alignments and found ok.

Run Mechanism check and found ok.

Engineer comments: instrument is ok.

Time Report

Category	Start	End	Hours	Invoice Type
05913616001-Service labour time	13/02/2023 8:30 PM	13/02/2023 9:39 PM	1.15	Contract

Field Service Report



Case number: CAS-0016640862 Order Number: WO-03246864 | Ser. Report Created Date:

Signature

Customer Signature

Mr. Ravi

Roche support Tadiparthi R Siddardha

Sign Here



Roche Signature

The customer acknowledges the service intervention as performed in accordance with Roche recommendations:

- This Service report has been signed by the authorized representative of your organization.
- Discrepancies if any must be intimated within a period of seven (7) days of the receipt of the email, in the absence of which the report will be final and conclusive.

- This report shall be governed by the laws of India and the courts at Mumbai shall have the exclusive jurisdiction to deal with the disputes arising hereunder.

Customer Support Center No: Toll Free No: 1800-123-7599 || Phone No: 044-43900345 ||

(Mon to Sat 8 AM to 8 PM)

e-mail: customersupportcenterIndia@roche.com ||||||||||||

Roche DiaLog: https://dialog.roche.com Service Report Version Number: 1