



Alere Medical Pvt. Ltd.
No. 404, BPTP Park Centra,
NH - 8, Sector - 30
Opp. 32nd Milestone, Gurugram
Haryana 122001, India
T: +91 124 4569000
F: +91 124 4569011

Calibration Certificate

Instrument Name : Urometer120
Serial Number : 12076624

Customer Name : Tesla Diagnostics Center
Address : karmanghat, Hyderabad

Standard used : Self Calibration Mode
Standard Traceability : In House Certificate

Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333 ± 10 .

QC Performance : Acceptable
Self Calibration : Acceptable

Performance Qualification : Passed

Calibration Date : 08.02.2023

Calibration : Passed

Calibration Valid up to : 07.02.2024

For Alere Medical Pvt Ltd.

V. U. P. T.

Upendrakumar V
Engineer Technical Services
M: 9052723178

