

ST-200Pro CALIBRATION CERTIFICATE

CUSTOMERS NAME: **Tesla Diagnostics**

ADDRESS: **Champapet Main Road, opp. Andhra Bank, Kharmanghat, Hyderabad.**

CONTACT NO:

SERIAL NO OF THE INSTRUMENT: **STP-01291**

CALIBRATION DATE: **01/10/2022**

VALID TILL: **30/10/2023**

CALIBRATION DONE BY: **Mr. KIRAN DATE**

Calibration Slopes:				
S.No.	Na ⁺	K ⁺	iCa	Cl ⁻
1	65.13	63.79	34.12	50.93
2	64.88	63.40	35.01	51.06
3	64.78	63.42	35.12	51.02
Mean	64.93	63.53	34.75	51.00
Ranges	35-100mV/dec	35-100 mV/dec	15 - 50 mV/dec	30 - 100 mV/dec
Status	O.K	O.K	O.K	O.K

Sensa Core Medical Instrumentation Pvt. Ltd.

Plot, No: 3, Export Promotion Industrial Park, Pashamylaram, Medak (Dist), 502307, T.S. INDIA.
Tel: +08455 - 223400, www.sensacore.com

ST-200Pro CALIBRATION CERTIFICATE

CUSTOMERS NAME: **Tesla Diagnostics**

ADDRESS: **Champapet Main Road, opp. Andhra Bank, Kharmanghat, Hyderabad.**

CONTACT NO:

SERIAL NO OF THE INSTRUMENT: **STP-01291**

CALIBRATION DATE: **01/10/2022**

VALID TILL: **30/10/2023**

CALIBRATION DONE BY: **Mr. KIRAN DATE**

Calibration Slopes:				
S.No.	Na ⁺	K ⁺	iCa	Cl ⁻
1	65.13	63.79	34.12	50.93
2	64.88	63.40	35.01	51.06
3	64.78	63.42	35.12	51.02
Mean	64.93	63.53	34.75	51.00
Ranges	35-100mV/dec	35-100 mV/dec	15 - 50 mV/dec	30 - 100 mV/dec
Status	O.K	O.K	O.K	O.K

Sensa Core Medical Instrumentation Pvt. Ltd.

Plot, No: 3, Export Promotion Industrial Park, Pashamylaram, Medak (Dist), 502307, T.S. INDIA.
Tel: +08455 - 223400, www.sensacore.com