



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	Sri Siva Health Care Unit
Address:	Karmanghat , Hyderabad
Province-Postal code	500079
Contact Name	Mr. Mustaffa

**Calibration Certificate**

Date	12/04/23 <small>(dd mmmm, yyyy)</small>	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DC3G804602
------	--	------------------	------------------	--------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

**High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3**

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Measuring cylinder	11.20/2118	21/11/2020

Specified Range@ 1.5ml/Minute	Measured Value
1.425 – 1.575 ml/minute	1.51

**Column Holder temperature Reference: Service Manual: Section 7.9**

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Fluke 179	43160402	16/11/2022

Specified Range @ 35°C	Measured Value
33.4°C – 36.4°C	35.0

**Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081**

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.021

The next calibration is due on	11/04/2024	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

*A. R. C. S. Prasad*