

CALIBRATION REPORT

Centrifuge Table top



Report No:	CAL-AR-2023-02-0033	Date:24-02-2023	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	FRU Ruksin CHC		
State	Arunachal Pradesh	District	East Siang

Medical Equipment Details			
ME Number	ARZ1EASCH0510045	Manufacturer	Unilab
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	NA
Department	General Lab-13 Room	Serial No	22757
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	15-02-2023	Due for Calibration	14-02-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%



Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0018	Digital Thermometer	METRAVI	DTM-900	150615975	25-08-2023

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	rpm	1500	1530	1425 to 1575	Pass
2	Speed	rpm	2500	2540	2375 to 2625	Pass
3	Speed	rpm	3500	3550	3325 to 3675	Pass

CALIBRATION REPORT

Centrifuge Table top



Conditions of Issue of report	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval.	
Callbration Status	Passed <input checked="" type="checkbox"/> Failed <input type="checkbox"/>
Calibrated By  Shashi Shekhar	Authorised Signatory  Rahul Kumar Kashyap