

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Report No:	CAL-AR-2023-02-0037	Date: 24-02-2023	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	FRU Ruksln CHC		
State	Arunachal Pradesh	District	East Slang

Medical Equipment Details			
ME Number	ARZ1EASCH0510050	Manufacturer	LG
ME Name	Refrigerators, Laboratory (Lab refrigerator)	Model	INTELL COOL
Department	Ictc Room-Opd Building	Serial No	810NLEK025348
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	15-02-2023	Due for Calibration	14-02-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0018	Digital Thermometer	METRAVI	DTM-900	150615975	25-08-2023

Performance Results						
No	Parameter	Set Temp	Unit	Measured Value	Acceptable Range	Remarks
1	Temperature	2°C	°C	2.1	1.5 to 2.5°C	Pass
2	Temperature	4°C	°C	4.3	3.5 to 4.5°C	Pass
3	Temperature	6°C	°C	6.2	5.5 to 6.5°C	Pass

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed

Failed

Calibrated By

A handwritten signature in black ink, appearing to be "Shashi Shekhar", written over a horizontal line.

Shashi Shekhar

Authorised Signatory

A handwritten signature in black ink, appearing to be "Rahul Kumar Kashyap", written over a horizontal line.

Rahul Kumar Kashyap