

**CALIBRATION CERTIFICATE-MINDRAY BS-240PRO**

This is to certify that Mindray BS 240PRO, fully automated biochemistry analyzer having serial no. BC-97000397 has been calibrated on 03.09.2022 and results are found well within range.

Lab Name	The Leprosy Mission Hospital, Barabanki (U.P.)
Instrument Name	Mindray BS-240PRO
Instrument Model/Sr No.	BS-240PRO / BC-97000397
Calibration Date	03.09.2022
Next Calibration Due Date	03.09.2023

Test Parameter	Required Value	Measured Value
Input Voltage	230~240 V	231
Incubation Temperature	37°C+/-1max	37°C
Reagent Tray Temperature	5~15°C	8°C
Detector Performance Check @All Wavelength	6000~16000	Within range
12V Supply	12+/-0.3Volts	12.1 Volts
5V Supply	5+/-0.3 Volts	5.06 Volts
24V Supply	24+/-0.3 Volts	24.2 Volts
Water Blank	(1800)of the water blank value	Within range



Calibrator Details:

MULTI SERA CALIBRATOR

Lot No.:150122002

Expiry: 31.10.2023

QC Details:

CLINCHEM MULTI CONTROL L1

Lot No.: 059322004

Expiry: 29.02.2024

QC Details:

CLINCHEM MULTI CONTROL L2

Lot No.: 059422004

Expiry: 29.02.2024

For Redcell Technocare Pvt. Ltd.

PRADEEP RAWAT
(Service Engineer)