

**Calibration Certificate Automated Hematology Analyzer**

This is to certify that 5 Part Hematology Analyzer BC 5150 having serial no. SR-73001780 has been calibrated on 01.09.2022 and results of 3 levels QC Low, Normal and High are well within range.

| | |
|---------------------------|--|
| Lab Name | The Leprosy Mission Hospital Barabanki (U.P.) |
| Instrument Name | 5-Part Hematology Analyzer BC-5150 |
| Instrument Model/Sr No. | BC-5150/ SR-73001780 |
| Calibration Date | 01.09.2022 |
| Next Calibration Due Date | 01.09.2023 |

| Background Check | Required Value | Measured Value |
|------------------|----------------|----------------|
| WBC | < 0.2 | 0.0 |
| RBC | < 0.02 | 0.0 |
| HCT | < 0.5% | 0.0 |
| Hb | < 0.1g/dL | 0.0 |
| PLT | < 10 | 5 |



Calibrator Details:

SC-CAL Plus

LOT No.: PLUS1221

Expiry: 10.11.20212

QC Details:

BC-5D

Lot No.: BC2209B

Expiry: 10.11.2022

For Redcell Technocare Pvt. Ltd.

PRADEEP RAWAT
(Service Engineer)

Redcell Technocare Pvt.Ltd.

B-116/8, Indira Nagar, Lucknow-226016 | Mob: 9415595860, 9793088855, E-mail: redcelltechnocare@gmail.com

