


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 230526A/N-119	<b>Page</b> : 1 of 1			
<b>CALIBRATED ON.</b> : 27 / 05 / 2023	<b>Ref. Std.</b> : CG-11			
<b>CALIBRATION DUE.</b> : 26 / 05 / 2024	<b>C.P. No.</b> : ACS/CP/07			
<b>SRF No. &amp; DATE</b> : 230526A/N Dt. 26.05.2023	<b>Phy. Inspection</b> : O.K.			
<b>ULR No. &amp; ISSUE DATE</b> : CC342723000000240F Dt. 30.05.2023	<b>Cal. At</b> : Site / Lab			
<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. Khasra No. 73 and 74 Situated at Moolraj Building Main Market Yusuf Sarai, New Delhi-110016				
<b>ITEM CALIBRATED</b> : Digital Temperature Indicator				
<b>Make</b> : RoHS	<b>Range</b> : (-50 to 110) °C			
<b>L. Count</b> :	<b>ID No.</b> : RCL/HEM/D.Temp-03			
<b>Location</b> : Haematology	<b>S. No.</b> : N.A			
<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 3°C } R.H. {40% to 60%}				
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	09/11/2023
<b>RESULTS</b> : ET Discipline				
<b>* U. C. INSTT. READS (°C)</b>	<b>STANDARD READING (°C)</b>	<b>ERROR (°C)</b>		
2.04	2	0.04		
4.04	4	0.04		
6.06	6	0.06		
8.10	8	0.10		
(average of readings)				

**REMARKS**

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 1.00^\circ\text{C}$  at coverage Factor  $k=1.96$ .

\*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

