


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230526A/N-120	Page	: 1 of 1
CALIBRATED ON.	: 27 / 05 / 2023	Ref. Std.	: CG-11
CALIBRATION DUE.	: 26 / 05 / 2024	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 230526A/N Dt. 26.05.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000241F Dt. 30.05.2023	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Khasra No. 73 and 74 Situated at Moolraj Building Main Market Yusuf Sarai, New Delhi-110016		
ITEM CALIBRATED	: Digital Temperature Indicator		
	Make : RoHS	Range	: (-50 to 110) °C
	L. Count :	ID No.	: RCL/HEM/D.Temp-04
	Location : Haematology	S. No.	: N.A
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
DUE DATE			
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846
			09/11/2023
RESULTS	: ET Discipline		
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.08	2	0.08	
4.10	4	0.10	
6.14	6	0.14	
8.18	8	0.18	
(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

