


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 230526A/N-101	<b>Page</b> : 1 of 1			
<b>CALIBRATED ON.</b> : 26 / 05 / 2023	<b>Ref. Std.</b> : ISO:8655:6			
<b>CALIBRATION DUE.</b> : 25 / 11 / 2023	<b>C.P. No.</b> : ACS/CP/08			
<b>SRF No. &amp; DATE</b> : 230526A/N Dt. 26.05.2023	<b>Phy. Inspection</b> : O.K.			
<b>ULR No. &amp; ISSUE DATE</b> : CC342723000000222F Dt. 30.05.2023	<b>Cal. At</b> : Site / Lab			
<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. Khasra No. 73 and 74 Situated at Moolraj Building Main Market Yusuf Sarai, New Delhi-110016				
<b>ITEM CALIBRATED</b> : Micro Pipette				
<b>Make</b> : Thermo Scientific	<b>Range</b> : (2 to 20) µl			
<b>L. Count</b> :	<b>ID No.</b> : RCL/HEM/PIP-01			
<b>Location</b> : Haematology	<b>S. No.</b> : TW08232			
<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 3°C } R.H. { 40% to 60% }				
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	04/01/2024
<b>RESULTS :</b> Mechanical Discipline				
<b>Nominal Capacity (µl)</b>	<b>Graduation (µl)</b>	<b>Nominal Value (µl)</b>	<b>Volume at 20 °C (µl)</b>	<b>Error (µl)</b>
20		10	10.07	(-) 0.07
		20	20.14	(-) 0.14

**REMARKS**

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

...End of Calibration Certificate....

