


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230526A/N-105	Page : 1 of 1
CALIBRATED ON. : 26 / 05 / 2023	Ref. Std. : ISO:8655:6
CALIBRATION DUE. : 25 / 11 / 2023	C.P. No. : ACS/CP/08
SRF No. & DATE : 230526A/N Dt. 26.05.2023	Phy. Inspection : O.K.
ULR No. & ISSUE DATE : CC342723000000226F Dt. 30.05.2023	Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
Khasra No. 73 and 74 Situated at Moolraj Building
Main Market Yusuf Sarai, New Delhi-110016

ITEM CALIBRATED : Micro Pipette	
Make : Thermo Scientific	Range : (10 to 100) µl
L. Count :	ID No. : RCL/BIO/PIP-05
Location : Clinical Biochemistry	S. No. : TW08013

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	04/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.10	(-) 0.10
100		50	50.18	(-) 0.18
		100	100.32	(-) 0.32

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

