



Calibration Certificate – Rx Imola

Instrument	Rx Imola Clinical Chemistry Analyzer
Serial No	7201-0816
Account Name	P. H. Medical Centre - Mumbai
Installation Date	10 May 22
Calibration Date	18 May 23
Next Calibration Due	19 Nov 23

This is to certify that this analyser has been inspected and calibrated for following parameters:

Test Parameter	Target Value	Obtained Value
Input Voltage	230 -240 V AC	231 V AC
Cuvette Check	<0.5 Abs for all cuvettes	<0.21 Abs
Photometer Check	<0.5 Abs for all wavelengths	<0.16 Abs
Incubator Temperature	37° C ± 0.3 max	37.0 ° C
Reagent Tray Temperature	8- 15 ° C	8.00 ° C
Detector Performance Check	8.5 to 9.5 Volts	8.7 to 9.00 Volts
12V Lamp Supply	12 ± 0.3 volts	12Volts
5V Supply	5 ± 0.3 volts	5.0 volts
24V Supply	24 ± 0.3 volts	24 volts

The results obtained are as per specifications & tolerance ranges. The above calibration was done with a MECO 801AUTO Digital Multimeter. (Calibration certificate is enclosed herewith)

For Randox Laboratories India Pvt Ltd

Pranav Dudwadkar.
Customer Support Engineer