

Biofusion PRO

CE ISO 9001-2015

STATUS PASSED

DEVICE ID : Micro Pipette

CALIBRATION REPORT

METHOD ID: VV 50

DESCRIPTION

V.V.Range-5-50 µl

ENVIRONMENTAL

FACTORS

DATE

02/05/2022

TEMP 20.00 C

BARO. PRESSURE

750.00mmHg

REL. HUMIDITY

50.00%

AIR DENSITY 0.001200g/cm3 Z FACTOR

1.002899/cm3/g

CUBIC EXP.

0.0000cm3/degC

SAMPLE

CALIBRATION STATISTICS

Vol.	No.	Wt.	Vol.	Vol.	Mean	Sd	Actual	Target	Status
5.0	1	4.98	4.98	5.00	5.01	0.20	2.29	0.10	passed
5.0	2	10.00	10.05	10.00	9.99	0.35	0.9	0.15	Passed
5.0	3	29.96	14.99	14.98	15.03	0.45	15.00	0.10	Passed
25.0	1	25.00	25.05						
25.0	2	49.85	24.88						
25.0	3	75.00	25.00						
50.0	1	50.00	49.75						
50.0	2	99.08	50.00						
50.0	3	149.85	49.99						

<
<
<

ality control by

Calibration & Final Q.C By

.....

.....

Signature

Signature

- ① Specifications conform to DIN 12650 Standards.
- ② Each Pipette is individually calibrated on Electronic Balance
- ③ Add.