



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3729

Distribution No.: 159-J

Month/Year: March/2023

Instrument ID: Haematology analyzer

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 01-06-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.1	5.1	10.2	10.36	0.038	-0.13	0	0.1	0.008	-0.76
RBC x10 <sup>6</sup> /µl	1	4.5	4.49	8.99	8.99	0.011	0.02	0.01	0.04	0.003	-0.67
Hb g/dl	1	14.8	14.8	29.6	29.3	0.029	0.40	0	0.1	0.008	-0.67
HCT%	1	43.2	43.1	86.3	88.55	0.234	-0.36	0.1	0.4	0.027	-0.67
MCV-fl	1	96.2	96.1	192.3	196.05	0.390	-0.32	0.1	0.3	0.023	-0.45
MCH-Pg	1	32.9	32.8	65.7	65.5	0.076	0.10	0.1	0.3	0.018	-0.67
MCHC-g/dl	1	34.3	34.2	68.5	65.9	0.163	0.59	0.1	0.3	0.020	-0.67
Plt. x10 <sup>3</sup> /µl	1	144	138	282	255	1.285	0.74	6	4	0.284	0.45
Retic %	2	30	28	58	24.65	0.520	2.34	2	1	0.062	0.90

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=2 , Poly=38 L=4, E=2, Mono/Promono=03 , B1=2 P.M.=8, Mye=15, Meta=28, Other=
RBC Morphology	3	Poly: 25 - 45, Myelo: 15 - 31, Meta: 10- 20, Lympho: 2- 7, Eosino: 1-4, Promyelo: 2-7, Blast: 1-4, Mono: 1 - 3, nRBC/Baso: 0-5
Diagnosis	3	Chronic Myeloid Leukemia (Chronic Phase)

## COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 159--J	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	308	302	83.44	86.75	2.98	5.96	13.58	7.29
RBC x10 <sup>6</sup> /µl	1	308	308	87.66	90.26	5.52	2.92	6.82	6.82
Hb g/dl	1	308	308	81.49	83.44	5.52	6.49	12.99	10.07
HCT%	1	308	304	91.12	89.8	5.92	3.62	2.96	6.58
MCV-f	1	308	304	95.72	92.43	1.97	2.96	2.31	4.61
MCH-Pg	1	308	304	88.82	89.14	6.58	2.63	4.6	8.23
MCHC-g/dl	1	308	304	93.42	89.14	3.29	6.58	3.29	4.28
Plt. x10 <sup>3</sup> /µl	1	308	304	94.41	91.12	3.29	6.25	2.3	2.63
ReticCount%	2	308	268	94.4	92.16	4.85	3.36	0.75	4.48
PS Assessment	3	308	268	Satisfactory :95.46%, Borderline Sat. :2.27%, Unsatisfactory :2.27%					

**Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3 \cdot SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x} - \bar{y}$ ) should be smaller than the check value ( $0.3 \cdot SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme


**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmalimseqap.com](http://www.ishtmalimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----