

ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Custom	ner name	M/S. RAI RISHI	Service request No. & date	P/02/28/04/23&28/04/2023
Aı		Master Para, Katwa S.D Hospital	ULR No.	CC223923000002404F
Add	ress	3 No Gate, Katwa, Dist-Purba	Cert. No.	AT/23000002404
		Bardhaman-713130, West Bengal	Date of Receipt of DUC	28/04/2023
			Date of calibration	28/04/2023
			Date of issue	29/04/2023
9		1	Suggested due date	27/04/2024

Instrument Details			
Instrument name	CENTRIFUGE	S.No.	
Make /Model No		Location	
Range / Size	0 to 3000 rpm	Accuracy	
Least Count	1 rpm	Visual Inspection	OK
I.D. No.	NEYA2 remi		

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	/A.C92792		
Certificate No.	QCTS/000246/10		
Calibration Validity	21/10/2023		
Calibration by	Quality Calibration Testing Sol.		

Environmental	Temperature	20±2°C	Calibration Reference	IS: 12508
Condition	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results

Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	100.9	100	
02.	202.4	203	
03.	503.7	505	
04.	1013	1010	±0.20 %rdg
05.	2065	2005	
06.	3011	3001	

Remarks:

(1) Standard equipment use for calibration are traceable to national/international standards.

(2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (MAHIPAL) Form No.- QF-47



Approved By (QM/TM)
(NEERALTYAGI)
Page No. 1 of 1