

ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S. RAI RISHI Master Para, Katwa S.D Hospital 3 No Gate, Katwa, Dist-Purba Bardhaman-713130, West Bengal	Service request No. & date ULR No. Cert. No. Date of Receipt of DUC	P/02/28/04/23&28/04/2023 CC223923000002406F AT/23000002406 28/04/2023
	Bardhaman-713130, West Bengal	Date of Receipt of DUC	
a e		Date of calibration	28/04/2023
		Date of issue	29/04/2023
		Suggested due date	27/04/2024

	Instru	ment Details	
Instrument name	Micro Pipette	Sr. No.	
Make / Model No.		Location	
Range / Size	10 to 100 μl	Accuracy	
Least Count	10 μl	Visual Inspection	OV.
I.D. No.	M-5349	v isual hispection	OK

	Detail of reference standar	ds & Major equipments used	
Equipment Name	A set of weight box	Digital Weighing Balance	_
Make	WEIGHTRONICS	AND	
Model / SR No.		GH-252/ 78002	
Certificate No.	TYCON/WB/12/274	TYCON/WB/02/23/327	
Calibration Validity	25/12/2023	21/02/2024	
Calibration by	Tycon Engineering	TYCON ENGINEERING	

Environmental Condition	Temperature	23± 3 °C		
	Relative Humidity		Calibration Procedure	
		THE STATE OF THE S	- The state of the	C1-33

Serial No.	Nominal Capacity in (µl)	alibration Results Measured Capacity in (μl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	10	9.98	±0.21 μl
02.	20	19.95	
02.	50	49.89	
03.	100	99.81	

Remarks:

(1) Standard equipment use for calibration are traceable to national/international standards.

(2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (MATHRAL) Form No.- OF-47



Approved By (QM/TM) (NEERAL TYAGI)
Page No. 1 of 1