

### Calibration Certificate

Instrument Name : Urometer 120  
Serial Number : 12211058  
Customer Name : Lifewell Diagnostics .P. Ltd  
Address : Punjab-Chandigarh  
Standard used : Self Calibration Mode  
Standard Traceability : In House Certificate

#### Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333 +10.

QC Performance : Acceptable

Self-Calibration : Acceptable


#### Performance Qualification : Passed

Calibration Date : 13.04.2023

Calibration : **Passed**

Calibration Valid up to : 12.04.2024

For **Abbott Diagnostics Medical Pvt. Ltd**



R.Muralidharan  
Engineer - Tech Services

