



Metropolis Healthcare Ltd., EQAS
TATA 1 MG LABS(TATA 1 MG TECHNOLOGIES PVT LTD) (803)
Outlier And Analyte Summary Report

Outlier Details For Cycle No 230104 and Sample No 01

Report Date : 19/05/2023

Analyte	Instrument	Result Value	Standard Unit
X Protein (Urine)	Clinitex status +	2+	-
X Blood (urine)	Clinitex status +	Trace	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ Delayed Result Entry

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Clinitek status +

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5.5	5,5,5,6
@ Specific Gravity	-	1.015	1.015,1.020,1.025
@ Glucose (urine)	-	3+	+2,+3,+4
X Protein (Urine)	-	2+	+1,Tr
@ Ketone (urine)	-	Negative	Neg
X Blood (urine)	-	Trace	Neg
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Negative	Neg
Legend @ : Acceptable		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	2
X : Unacceptable		Evaluated Parameters	8
# : Not Evaluated		Outlier Parameters (X)	2
⊙ : Delayed Result Entry		EQAS Score Clinical Path - Urine Routine Chemistry	75.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070