

CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Report No:	CAL-JK-Z2-2022-06-1600	Date: 01-06-2022	Page 01 of 02
------------	------------------------	------------------	---------------

Customer Details			
Health Facility	Srinagar (JLNM) DH		
State	Jammu and Kashmir	District	Srinagar

Medical Equipment Details			
ME Number	JKZ2SRIDH0200106	Manufacturer	Remi
ME Name	Centrifuges, Table top (Centrifuge)	Model	R-8C
Department	ICTC Section	Serial No	ZCLN-37164
Calibrated at	On site	Condition of ME	Functioning Good
Date of Calibration	11-05-2023	Due for Calibration	10-05-2024

Environmental Conditions			
Temperature in °C	22°C	Humidity in % Rh	40




Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0263	Tachometer	Amprobe	TACH-10	19110026	04-01-2024

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	Rpm	1000	999	925 to 1075	Pass
2	Speed	Rpm	2000	2021	1925 to 2075	Pass
3	Speed	Rpm	3000	3015	2925 to 3075	Pass

CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Conditions of Issue of report	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval.	
Calibration Status	Passed <input checked="" type="checkbox"/> Failed
Calibrated By:  Farhan Farooq (BME)	Authorised Signatory:   Aman Bali (Regional Manager)