



LifeWell

PT EQAS RESULT REVIEW FORM

Issued by QA: Amey 22/05/2023

Department: <u>Clinical pathology</u>	Month / Year: <u>May 2023</u>
PT Provider: <u>Metropolis Healthcare Ltd</u>	Distribution / Lot No: <u>230104</u>
Analyte(Test): <u>PH, specific gravity, glucose, protein, ketone, Blood, Bilirubin, Urobilinogen Nitrite, crystal.</u>	
Machine used: <u>Manual</u>	Kit used: <u>Abbott</u>
Test done by: <u>Abhishek</u>	Done on: <u>11/05/2023</u>
Report uploaded by: <u>Abhishek</u>	Report uploaded on: <u>12/05/2023</u>
Software used: <u>online web portal</u>	
EQAS Report received on <u>19/05/2023</u>	
Reviewed by: <u>Abhishek 19/05/2023</u>	Approved by: <u>Amey 19/05/2023</u>
Observations: <u>All Result found satisfactory</u>	

\*If results are ungraded, please attach self-assessment. If satisfactory no corrective and preventive action required. But if unsatisfactory, corrective and preventive action is required with optimum evidence.

Self-Assessment Result:  Satisfactory  Unsatisfactory

Attachment:  Yes  No

Please investigate failure for unsatisfactory performance and submit a report to QA and Medical Director in N/A Days.


Approved by

Signature (Lab Director/Designee) [Signature] 22/05/2023

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated 🕒 Delayed Result Entry			

Problem Classification: there is no out layer

Corrective Action: no action Required

Reviewed by: 

Dated: 24/05/2023

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5,5,5,6
@ Specific Gravity	-	1.025	1.015,1.020,1.025
@ Glucose (urine)	-	3+	+2,+3,+4
@ Protein (Urine)	-	Trace	+1,Tr
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	Negative	Neg
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
Ⓢ : Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	0
Ⓞ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ Delayed Result Entry			


Problem Classification:

*There is no outlier*

Corrective Action:

*No action Required*

Reviewed by:



Dated:

*22/05/2023*

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Struvite crystals	Struvite crystals
Legend @ : Acceptable @ : Acceptable X : Unacceptable # : Not Evaluated Ⓞ : Delayed Result Entry			
		Total Parameters	1
		Not Evaluated Parameters	0
		Evaluated Parameters	1
		Outlier Parameters (X)	0
		EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070